

08.12.2020

Corrigendum 01

With reference to the tender notification for “**Bioreactor**” **Tender Notification No:** NITK/CH/DST-FIST-Bioreactor/2020-2021/316 dated: 03.12.2020, the following specifications are amended with reference to Annexure H of the tender document.

1. Instead of” Top Driven PMDC brushless motor, Glass vessel (In-situ sterilizable) with Glass Jacketed”, new specification should be "AC Servo motor bottom driven with mechanical seal, Glass vessel (In-situ sterilizable)
2. Under "Temperature control", this specification, "Direct immersed electrical heater 400W x2nos usable for minimum working volume and cooling coil to maintain the temperature for the temperature control" is added.
3. Instead of "PH Measurement and control: There should be provision for accommodating PH probes of Mattler make or better."... new specification "PH Measurement and control: Insitu Sterilizable PH probe of Mettler toledo make or better with the transmitter"
4. Under DO measurement and control: Spelling mistake in the word "Mattler". The correct spelling is "Mettler toledo"

Sd/-
Prepared by Buyer

Sd/-
Approved by
HOD, Chemical Dept NITK, SURATHKAL

Date of updation: 08.12.2020

Seal of the Office of the TIA